



# Certificate

## Trace Oxygenates In Gasoline

LOT #051608CC

Expiration Date : May 2011

### CALIBRATION CURVE LEVELS IN ISO-OCTANE

	STD#1 ppm wt	STD#2 ppm wt	STD#3 ppm wt	STD#4 ppm wt	STD#5 ppm wt	STD#6 ppm wt
ETBE	18	88	176	440	880	1759
MTBE	17	86	172	429	859	1717
DIPE	17	85	170	426	852	1703
TAME	18	92	184	461	921	1842
Metanol	19	94	188	470	940	1881
n-Propanol	19	95	190	476	952	1904
I-Propanol	19	95	189	473	945	1891
t-Butanol	19	96	191	478	956	1913
s-Butanol	19	97	194	485	971	1941
I-Butanol	19	97	194	486	972	1944
n-Butanol	20	98	195	488	976	1952
t-Pentanol	20	98	196	490	979	1958
1,2 DME Int.Standard	313	313	313	313	313	313
Ethanol	15%	15%	15%	15%	15%	15%

\* = combined concentration of indicated components

-STANDARD PREPARATION COMPLIES WITH ISO 17025

## Separation Systems, Inc.

Experts in Gas Chromatography

Gulf Breeze, Florida, 32561

Phone 850-932-1433 www.separationsystems.com

KEEP REFRIGERATED

KEEP REFRIGERATED